

OXYGEN MEASUREMENT: min. Hp=50mm

L = 120 + Vz + Vh + Hp - 135

pH,ORP,CONDUCTIVITY MEASUREMENT : min. Hp=65mm

L = 120 + Vz + Vh + Hp - 195

Interconnection cable for SPO sensors we recommend to extend for length equal to the length of the probe plus 100 cm, to have possibility to move probe during calibration and cleaning.

SPO 41 Submersible probe – connecting rod length