


ANALOGUE OUTPUT SIGNALS

OUT.CLAMP	OUTPUT	
12 13	1	ClO ₂ 1
14 15	2	ClO ₂ 2 / pH
16 17	3	AR.AVERAGE/CONTROLLER 2
18 19	4	CONTROLLER 1


ANALYZER-PROBE CONNECTIONS, CHLORINE DIOXIDE AND pH MEASUREMENT

INPUT.CLAMP	MPC66	PROBE		
20	INPUT 1	20	ClO ₂ 1	PROBE SPR42 , PK41
21	INPUT 1	19	ClO ₂ 1	
22	INPUT 2	18	pH	
23	INPUT 2	17	pH	

CHLORINE DIOXIDE MEASUREMENT (WITHOUT pH)



CHLORINE DIOXIDE AND pH MEASUREMENT (TEMPERATURE USED FOR COMPENSATION ONLY)



z	GREEN
c	BLACK
b	WHITE
r	RED
me	MAIN ELECTRODE
re	REFERENCE ELECTRODE